Serial No. C 135023

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com

WEIGHTRON BILANCIAI
INDUSTRIAL WEIGHING SPECIALISTS
BILANCIAI GROUP

Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com

UKWF
Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Sherrington Waste LTD
Digbeth
LOCATION Birmingham

Machine Details

Cust ref	weighbridge
Make	BILANCIAI
Model	DD 700
Ser No	25100668
Capacity	140 1000kg
e/d	120 1kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
WBS0-71		13-10-25
FC	TE12891	28-5-25

Cert Interval	12 months	
Next Cert	OCT	2026

Tolerance +/-	kg
0- 10000	20
10000 - 40000	40

Comments ..... Make up
..... 5320 kg

2	3
1	4

4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
30000	30000	0	30000	0
40020	40020	20		

Repeatability

Load	Ind	Ind	Ind

Repeatability 31120 kg

Load	Ind	Ind	Ind
31120	31120	31120	31120

Eccentric Load

Reading		Dev	Reading		Dev
1			2		
3			4		
5			6		
7			8		

Eccentric Load 13000 kg

Reading		Dev	Reading		Dev
1	13000	0	2	13000	0
3	13000	0	4	13000	0
5			6		
7			8		

Signature ..... Name ..G. Pritchard..... Date ..14-10-25.....

C1.3.88.1N